



Clonal Technologies

For more information
contact Clonal Technologies
on 1300 795 335

See our range online at www.clonaltech.com.au

Email: john.dubois@clonaltech.com.au

33 Fulcrum Street
Richlands Queensland 4077

Detectabuse® Urine Liquid Control

Detectabuse® urine controls are designed to monitor and validate the performance of drugs of abuse detection devices in accordance with Australian Standard AS/NZ 4308:2008.



Features

- Available in -50% and +50% cutoffs
- Each control bottle contains 20ml of stabilized human urine
- Only 2ml of control solution is needed to conduct a control test
- Positive control urines have been spiked with authentic reference drug standards and/or appropriate metabolites
- Negative control urines are certified negative
- Once opened, bottle can be resealed and either refrigerated or refrozen

Benefits

- Allows positive and negative controls to be conducted to Australian Standard levels
- Specific gravity, pH and creatinine fall within the limits of normal human urine
- One control bottle allows for 10 control tests to be conducted
- Laboratory certified control solution levels
- Laboratory certified control solution levels
- No wastage and will remain stable for up to six months

UZ-18001153 - Detectabuse® Urine Control Solution Negative (-50%) Cutoff (AS/NZ 4308:2008) Single 20ml Bottle

UZ-18001154 - Detectabuse® Urine Control Solution Positive (+50%) Cutoff (AS/NZ 4308:2008) Single 20ml Bottle